

Display Report

Analysis Info

Analysis Name D:\Data\ZHL\MS\2026\MS-02Y\MS-20260204\20260123_P1-E-7_01_1378.d
Method tune_pos_standard.m
Sample Name 20260123
Comment

Acquisition Date 2/4/2026 10:32:10 AM
Operator BDAL@DE
Instrument maXis impact 8282001.00129

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.5 Bar
Focus	Active	Set Capillary	3500 V	Set Dry Heater	180 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	4.0 l/min
Scan End	1500 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Source
		Set Corona	0 nA	Set APCI Heater	0 °C

